



CERTIFICATE OF COMPLIANCE

Certificate Number: SGSNA/21/CA/00008X

Contract Number: 801086

Certificate Project Number: SUW-CERT210200014

Certified Product: Compression Pressure Sensor

Trademarks: IMES; Windrock

Model(s): 6330, 6331, 6332, 6333, 6334, 6335, 6336, 6337, 6338, 6339, 02060090 R 1.0, 02060100 R 1.0, 02060110 R 1.0, 02060120 R 1.0, 02060130 R 1.0, 02060140 R 1.0, 02060120 R 2.0, 02060130 R 2.0, 02060140 R 2.0, 02060150 R 1.0

Technical Data: Input: 9-28Vdc, 100mA, Output 4-20mA

Class I, Division 1, Groups A-D, T4, Ta: -40°C to +121°C IP 67

Vmax = 28V, Imax = 100mA, Ci = 59nF, Li = 132µH,
when installed as per the installation drawing 920038

Class I, Division 2 Groups A-D, T4 Ta: -40°C to +121°C IP 67

Pin A: Voc = 24V, Isc = 100mA, Ca = 0.46µF, La = 200µH

Pin D: Vmax = 28V, Imax = 100mA, Ci = 59nF, Li = 132µH
when installed as per the installation drawing 920038

Certificate Holder: IMES GmbH

Dr.-Herbert-Kittel-Str. 2, 87600, Kaufbeuren, Germany

This certificate supercedes previous certificates issued with the same certificate number. Certification is valid when products are indicated on the SGS directory of certified products at www.sgs.com or using the QR code below. The product is certified according to ISO/IEC Guide 17067, Conformity assessment - Fundamentals of product certification, System 3, and in accordance with:

UL 913, 8th Edition, Rev. Dec. 6, 2019

UL 60079-0, 7th Ed. Rev. March 26, 2019

CAN/CSA C22.2 No. 60079-0:19

UL 60079-11, 6th Ed. 02/15/2013

CAN/CSA C22.2 No. 60079-11:14 Rev. February 2014

Conditions of Acceptability:

Must be installed per control drawing 920038.

Authorized by:



Paul Krauss
Certifier

Effective date: 24 April 2021



Certification Body

Consumer and Retail Services, a division of SGS North America Inc.
620 Old Peachtree Road, Ste. 100, Suwanee, GA 30024, USA
t +1 770 570 1800 f +1 770 277 1240 www.sgs.com

